



PHENOL RED KIT

OCTA-SLIDE 2, pH 6.8-8.2

CODE 2110-01


QUANTITY	CONTENTS	CODE
30 mL	Phenol Red Indicator	2211-G
2	Test Tubes, 2.5-5-10 mL, plastic, w/caps	0106
1	Phenol Red Octa-Slide 2 Bar, 6.8-8.2	2185-01
1	Octa-Slide 2 Viewer	1101

*WARNING: Reagents marked with an * are considered to be potential health hazards. To view or print a Safety Data Sheet (SDS) for these reagents go to www.lamotte.com. Search for the four digit reagent code number listed in the contents list. Omit any letter that follows or precedes the four digit code number. For example, if the code is 4450WT-H, search 4450. To obtain a printed copy, contact LaMotte by e-mail, phone or fax.

Emergency information for all LaMotte reagents is available from Chem-Tel (US, 1-800-255-3924) (International, call collect, 813-248-0585).

To order individual reagents or test kit components, use the specified code number.

USE OF THE OCTA-SLIDE 2 VIEWER




The Octa-Slide 2 Viewer should be held so non-direct light enters through the back of the Viewer. Insert the Octa-Slide 2 Bar into the Viewer. Insert the reacted sample into the top of the Viewer. Match the color of the reaction to the color standards.

Warning! This set contains chemicals that may be harmful if misused. Read cautions on individual containers carefully. Not to be used by children except under adult supervision.


PROCEDURE

1



Insert Phenol Red Octa-Slide 2 Bar (2185-01) into the Octa-Slide 2 Viewer (1101)

2




Fill a test tube (0106) to the 10 mL line with sample water.

3




Add 8 drops of Phenol Red Indicator (2211).

4




Cap and mix.

5



Insert test tube into Octa-Slide 2 Viewer (1101).

6



Match sample color to a color standard. Record as pH.

LaMOTTE COMPANY

Helping People Solve Analytical Challenges

802 Washington Ave. · Chestertown · Maryland · 21620 · USA

800-344-3100 · 410-778-3100 [Outside U.S.A.] · Fax: 410-778-6394

Visit us on the web at www.lamotte.com